

A SEmi-Adiabatic Matched-Phase Spin Echo (SEAMS) PINS Pulse-Pair for B_1 -Insensitive Simultaneous Multislice Imaging

Rebecca E. Feldman,^{1*} Haisam M. Islam,² Junqian Xu,^{1,3} and Priti Balchandani¹

Purpose: Simultaneous multislice (SMS) imaging is a powerful technique that can reduce image acquisition time for anatomical, functional, and diffusion weighted magnetic resonance imaging. At higher magnetic fields, such as 7 Tesla, increased radiofrequency (RF) field inhomogeneity, power deposition, and changes in relaxation parameters make SMS spin echo imaging challenging. We designed an adiabatic 180° Power Independent of Number of Slices (PINS) pulse and a matched-phase 90° PINS pulse to generate a SEmi-Adiabatic Matched-phase Spin echo (SEAMS) PINS sequence to address these issues.

Methods: We used the adiabatic Shinnar Le-Roux (SLR) algorithm to generate a 180° pulse. The SLR polynomials for the 180° pulse were then used to create a matched-phase 90° pulse. The pulses were sub-sampled to produce a SEAMS PINS pulse-pair and the performance of this pulse-pair was validated in phantoms and in vivo.

Results: Simulations as well as phantom and in vivo results, demonstrate multislice capability and improved B_1 -insensitivity of the SEAMS PINS pulse-pair when operating at RF amplitudes of up to 40% above adiabatic threshold.

Conclusion: The SEAMS PINS approach presented here achieves multislice spin echo profiles with improved B_1 -insensitivity when compared with a conventional spin echo. **Magn Reson Med** 000:000–000, 2015. © 2015 Wiley Periodicals, Inc.

Key words: spin echo; simultaneous multi-slice; power independent of number of slices; PINS; SEmi-Adiabatic Matched-Phase Spin Echo; SEAMS; ultrahigh field; MRI

INTRODUCTION

Magnetic resonance imaging (MRI) at high magnetic fields, such as 7 Tesla (7T), offers increased signal-to-noise ratio (SNR) and enhanced contrast when compared with

conventional field strengths (1,2). This increased SNR can be parlayed into unprecedented spatial resolution or faster acquisitions for structural, functional, and spectroscopic imaging sequences (1–4). Spin echo preparations are often used for diffusion-weighted imaging (5–8) and may also be useful for functional MRI (9,10) at high magnetic fields. However, spin echo imaging at 7T faces several challenges that must be overcome to capitalize on the potential advantages. Three of these issues are: (i) increased inhomogeneity in the applied radiofrequency (RF) field (B_1 field) leading to nonuniformities in signal and contrast when using conventional RF pulses (11); (ii) increased RF power deposition in tissue, as measured by the specific absorption rate (SAR), limiting the strength and number of RF pulses in sequences (11); and (iii) changes in characteristic T_1 and T_2 tissue relaxation parameters that affect sequence timing and image contrast (2,12,13).

Simultaneous multislice (SMS) imaging using multi-band (MB) excitation and parallel imaging is a powerful technique which may be applied to reduce MR image acquisition time (14–18). SMS imaging is achieved by simultaneously exciting several slices in the region of interest (12–15). With the use of multiple receiver coil channels, the sensitivity profiles of the coils are used to disentangle the slices during reconstruction (18,19). One method of creating multislice excitations uses RF pulses with slice profiles that are spatially shifted off-center through a sinusoidal modulation of the wave form. The complex summation of these individual component RF pulses, with shifted spatial slice profiles, results in a single composite MB RF pulse. However, the RF power deposition of this type of MB pulse increases with the number of slices (17). Due to the quadratic increase of power deposition with field strength, the number of slices achievable by the conventional SMS technique at 7T is restricted by transmitter B_1 peak power and SAR safety limits. Furthermore, conventional MB composite pulses, particularly 180° MB pulses used in spin echo SMS sequences, are susceptible to slice amplitude and profile attenuation due to the significant B_1 field variation at 7T.

An alternative multislice pulse design method, the Power Independent of the Number of Slices (PINS) technique, may be used to produce pulses that excite multiple discrete slices simultaneously without increasing power deposition above that required for a single slice excitation (20). This is accomplished by interleaving the single-slice RF waveform, modulated with a comb function, with a gradient pulse train. The RF pulse used as the basis for the PINS pulse can be designed to provide the desired bandwidth (BW) and duration. Thus, specialized RF pulses, designed to address the technical challenges of

¹Translational and Molecular Imaging Institution, Icahn School of Medicine at Mount Sinai, New York, New York, USA.

²Bioengineering Department, Stanford University, Stanford, California, USA.

³Center for Magnetic Resonance Research, University of Minnesota, Minneapolis, Minnesota, USA.

Grant sponsor: NIH-NINDS; Grant number: R00 NS070821; Grant sponsor: Icahn School of Medicine Capital Campaign; Grant sponsor: Translational and Molecular Imaging Institute and Department of Radiology; Grant sponsor: Icahn School of Medicine at Mount Sinai; Grant sponsor: Siemens Healthcare.

*Correspondence to: Rebecca Feldman, Ph.D., Translational and Molecular Imaging Institution, Icahn School of Medicine at Mount Sinai, 1470 Madison Avenue, New York, NY 10029. E-mail: rebecca.feldman2@mountsinai.org, Twitter: @TMInyc

Received 13 November 2014; revised 13 January 2015; accepted 21 January 2015

DOI 10.1002/mrm.25654

Published online 00 Month 2015 in Wiley Online Library (wileyonlinelibrary.com).

© 2015 Wiley Periodicals, Inc.

ultrahigh field MRI, may be transformed into multislice pulses through the application of the PINS technique. Previous work has applied the PINS technique to adiabatic VERSE-DANTE pulses to help combat the effect of B_1 inhomogeneity (21). However, adiabatic pulses generate nonlinear phase across the selected slice profile and, in a typical spin echo sequence, a linear phase across the slice can only be achieved through the application of a second, identical adiabatic RF pulse. This twice-refocused approach permits the use of adiabatic pulses in a spin echo sequence, but at the cost of increased echo time (TE) and SAR. It has been shown that it is possible to design an excitation pulse to generate a complementary phase profile across the slice that corrects the phase deposited by a quadratic-phase adiabatic 180° pulse (22–24). Combining these two RF pulses results in a semiadiabatic excitation-refocusing pulse-pair that can be used for spin echo based imaging applications. Conversion to a multislice pulse-pair is then possible through application of the PINS technique.

In this work, we designed, implemented, and validated an adiabatic Shinnar Le-Roux (SLR) 180° PINS pulse and a matched-phase 90° PINS pulse to generate a multislice spin echo with refocused quadratic phase and improved immunity to B_1 -inhomogeneity. The pulse-pair was used as the basis of a SEmi-Adiabatic Matched-phase Spin echo (SEAMS) PINS sequence, which may be applied to accelerate spin echo sequences, especially when imaging at high fields.

METHODS

Pulse-pair Design and Implementation

Figure 1 is an overview of the design path used to develop a SEAMS PINS sequence. We used the adiabatic SLR algorithm (25,26) to create a 180° pulse with a bandwidth of 1.11 kHz, a duration of 7 ms and peak RF amplitude of $17 \mu\text{T}$. As previously described in Balchandani et al (26), the frequency profile of the 180° pulse design was the response of a least-squares linear-phase filter set to have the desired bandwidth. The filter design was done using the “firls” function in MATLAB (The Mathworks Inc., Natick, MA) $f_{\text{FIR}} = \text{firls}(N, F, AW)$ where N is the number of samples, F is a vector of frequency band edges given in the range $[0, \pi]$ but normalized to $[0, 1]$, A is a vector that specifies the desired amplitude of the frequency response of the filter, and W contains the relative ripple amplitudes of the pass and stop bands. For our design, $N = 355$, $F = \frac{1}{\pi}[0 \ 0.005\pi \ 0.025\pi \ \pi]$, $A = [1 \ 1 \ 0 \ 0]$ (a low pass filter), and $W = [0.1/8 \ \sqrt{0.01}/2]$. To introduce adiabatic behavior and distribute RF energy more uniformly, 2500 cycles of quadratic phase was applied across the frequency response for the filter. The $B_{180}(z)$ polynomial was then calculated as the Fourier transform of the resultant frequency profile and a minimum-phase $A_{180}(z)$ polynomial was calculated from $B_{180}(z)$. The $A_{180}(z)$ and $B_{180}(z)$ polynomials were used as inputs for the inverse SLR transform to produce an RF pulse waveform which was truncated, as described by Balchandani et al (26), to the portion of the result of the inverse SLR transform over which 98% of the RF power is deposited. This produced the desired pulse duration of 7 ms.




FIG. 1. Schematic diagram showing design flow of a SEAMS PINS pulse-pair.

In the second step, we designed the matched-phase 90° pulse as described in previous work (22). Note that this 90° pulse compensates the nonlinear quadratic phase deposited by the 180° pulse, but is inherently not adiabatic. Because high flip-angle RF pulses, such as 180° pulses, are more sensitive to B_1 variation than lower flip-angle excitation pulses, using an adiabatic 180° pulse with a matched nonadiabatic 90° pulse still greatly diminishes the deleterious effects of B_1 inhomogeneity on the spin echo signal (22). To obtain complementary phase in the excitation pulse, the $B_{90}(z)$ polynomial was calculated from the $B_{180}(z)$ polynomial using Eq. [1].

$$B_{90} = \frac{1}{\sqrt{2}} B_{180}^2 \quad [1]$$

The $A_{90}(z)$ polynomial was then calculated from $B_{90}(z)$ and both polynomials were set as inputs to the inverse SLR transform to generate the matched-phase 90° pulse. We set the pulse duration of the 90° pulse to be equivalent to the 180° pulse, as described by Park and Garwood (23) in condition II for phase-matching. Next, the resulting matched-phased semiadiabatic pulse-pair was transformed into a PINS pulse-pair.

The adiabatic PINS 180° refocusing pulse was created by sampling the adiabatic SLR 180° pulse with a comb function, creating evenly spaced nulls in the pulse waveform interleaved between RF pulse samples or lobes. The number of lobes in the 180° pulse (n) was calculated using Eq. [2] (21) to achieve a ratio between slice thickness (Th) and slice separation (S) of $\sim 1/8$, given the refocusing pulse bandwidth (BW) and RF pulse duration (t_{rf}).

$$n = BW \times t_{rf} \times \left(\frac{S}{Th} \right) \quad [2]$$

Figure 2 illustrates the resulting magnitude (Fig. 2D) and phase (Fig. 2E) of a 14 ms ($t_{rf} = 7$ ms), 62-lobe, 180° pulse. The excitation pulse was sampled to create the same slice separation and thickness ratio as the refocusing pulse. This resulted in a 90° pulse with 124 equally spaced lobes, magnitude and phase shown in Figures 2A




FIG. 2. SEAMS PINS pulse-pair. Amplitude of the SEAMS PINS RF waveform for matched-phase excitation (A) and refocusing (D). Associated phase for the excitation (B) and refocusing (E) pulses. C, F: Gradient waveforms timed to coincide with the nulls in the RF pulse. The insets in A, B, and C show the details of a 1 ms section of the waveforms.

and 2B. In our particular implementation, the total time (including the gradient lobes) of the excitation pulse was designed to be 50% longer than the total duration of the spin echo refocusing pulse to adhere to peripheral nerve stimulation (PNS) system limits. The excitation pulse consists of twice as many lobes as the refocusing pulse, however each RF lobe is 1/2 the duration, resulting in a total of 7 ms of RF sampling for both pulses. The RF nulls are the same duration for both the excitation and the refocusing pulse; thus the gradient lobe amplitude (Fig. 2C) played with the excitation pulse were identical to the gradient lobes used during the refocusing pulse (Fig. 2F). We could not shorten the gradient lobes any further without exceeding PNS limits on our system. The final excitation pulse design was 21 ms in total duration including 14 ms of null time to play gradients and 7 ms of RF sampling.

The gradient pulse trains, for both the excitation (Fig. 2C) and refocusing (Fig. 2F) pulses, were designed to interleave within the nulls of the PINS pulse. The time-integral of each gradient lobe (A) is given in Eq. [3] where $G(t)$ is the gradient amplitude and τ is the duration of the gradient lobe. To maximize A for a fixed small τ and typical maximum gradient slew rate on human scanners, each gradient lobe was designed to be triangular, with a maximum amplitude G_{amp} .

$$A = \int_{t_0}^{t_0+\tau} G(t) dt = \frac{1}{2} G_{amp} \tau \quad [3]$$

Thus, the separation between the slices can be calculated using Eq. [4] where γ is the gyromagnetic ratio.

$$S = \frac{1}{\frac{\gamma}{2\pi} A} = \frac{2}{\frac{\gamma}{2\pi} G_{amp} \tau} \quad [4]$$

Final Pulse Sequence and Reconstruction

The completed SEAMS PINS pulse-pair and associated gradient train were inserted into a spin echo pulse

sequence, replacing the conventional excitation and refocusing pulses, to create the SEAMS PINS sequence. We validated the performance of the pulse-pair for spin echo imaging in phantoms and in humans.

To reconstruct the aliased images, a set of weighting factors, corresponding to the individual coil sensitivities for each slice position, had to be calculated. To calculate this weighting, a low resolution (64×64) data set was acquired over the entire volume. For reconstruction, the raw data from each coil was retrieved and combined with the weighting factors, and a slice-GRAPPA MATLAB (The Mathworks Inc., Natick, MA) script (19,27) was modified to calculate the weights and un-alias the overlapping signals into individual separated slices.

Simulations

The $A_{PINS}(z)$ and $B_{PINS}(z)$ polynomials were used to simulate the final SEAMS PINS pulse-pair spectral profile. When the initial longitudinal magnetization is assumed to be 1, Eqs. [5–7] are equivalent to the calculations performed by a discrete-time Bloch simulator (25).

$$M_{xy} = 2 \text{conj}(A_{PINS90}(z)) * B_{PINS90}(z) \quad [5]$$

$$M_{ref} = B_{PINS180}(z) * B_{PINS180}(z) \quad [6]$$

$$M_{fin} = \text{conj}(M_{xy}) M_{ref} \quad [7]$$

Where M_{xy} is the transverse magnetization after the initial excitation by the matched-phase 90° PINS pulse and M_{ref} is the refocusing profile for the adiabatic PINS 180° pulse. The final magnetization (M_{fin}) at the spin echo can be calculated as the effect of the 180° pulse on the transverse magnetization produced by the 90° pulse (Eq. [7]) (25). The multislice magnitude and phase profiles of the adiabatic PINS refocusing pulse, obtained using Eq. [6], are shown in Figures 3A and 3C. The magnitude and phase of the multislice spin echo profile obtained using Eq. [7] are shown in Figures 3B and 3D. The nearly flat phase obtained across the slices in the final spin echo, after the application of the linear refocusing slice




FIG. 3. Slice Profiles for 180° adiabatic PINS pulse and final spin echo at three different slice positions. The amplitude (A) and phase of multislice (C) profile produced by the 180° adiabatic PINS pulse and the amplitude (B) and phase of spin echo (D) generated by the matched-phase adiabatic PINS pulse-pair. C: Quadratic phase deposited across the slices by the 180° adiabatic PINS pulse is largely refocused in (D) the spin echo, due to compensation by the matched-phase excitation pulse.

gradient, demonstrates that the quadratic phase is largely refocused in the final spin echo. This phase-matching obviates the need for a second adiabatic refocusing pulse that might otherwise be required to cancel nonlinear phase.

The B_1 -insensitivity of the SEAMS PINS pulse-pair was investigated by simulating the spin echo profile (using Eq. [7]) for RF overdrive factors ranging from 0.5 to 1.5, in increments of 0.1. The RF overdrive factor (ODF), given in Eq. [8], is a way of measuring the sequence response to a degree of B_1 -inhomogeneity. ODF is equal to the RF amplitude of the applied pulse (P_{app}) divided by the RF amplitude of the pulse at adiabatic threshold (P_{thresh})

$$\text{ODF} = \frac{P_{app}}{P_{thresh}} \quad [8]$$

Therefore, an ODF of 1.5 would mean a 50% increase of pulse amplitude above the adiabatic threshold. Maximum magnetization for the center slice was compared with a conventional nonadiabatic spin echo pulse-pair for the same range of ODFs. In the case of the conventional spin echo, ODF was calculated as the ratio of the applied RF pulse amplitude to the nominal RF pulse amplitude (i.e., the amplitude at which the pulse was designed to operate to achieve a flip angle of 180°).

Phantom Experiments

The performance of the SEAMS PINS pulse-pair was verified in a cylindrical water phantom (water bottle) and a spherical water phantom. All experiments were performed on a 7 Tesla (T) actively shielded whole body MRI scanner (Siemens MAGNETOM 7T, Siemens, Erlangen), equipped with a SC72CD gradient coil ($G_{max} = 70$ mT/m and max slew rate = 200 T/m/s), using a single channel transmitter and a 32-channel receive head coil (Nova Medical, Wilmington, MA). Pulse sequence parameters for the phantom experiments are summarized in Table 1.

The slice-selectivity and B_1 -insensitivity of the pulse sequence were tested on a cylindrical water phantom (created using a water bottle) by applying the SEAMS PINS sequence with the readout gradient played along the slice-select (axial) direction. This made it possible to obtain a projection through the multislice profile. We

then applied the SEAMS PINS sequence with the RF pulses powered at $\text{ODF} = 0.5\text{--}1.5$ in steps of 0.1 to determine whether the measured B_1 -insensitivity of the multislice profile agreed with simulated behavior.

Phantom experiments on a spherical water phantom were conducted to demonstrate in-plane slice behavior for a range of B_1 values as well as successful separation of overlapping slices. The SEAMS PINS sequence was powered with $\text{ODF} = 0.5, 1.0, 1.1, 1.2, 1.3, 1.4, 1.5,$ and 2.0. For this experiment, the readout gradient was played out in the conventional way, perpendicular to both the phase-encode and the slice-select gradients. The resulting aliased images were separated using the weightings calculated from a low resolution reference scan of the volume. A 10.4 mT/m gradient was used with a SEAMS PINS sequence designed to obtain a 1/8 slice separation ratio for 5 mm thick slices, resulting in a slice separation of 40 mm. In the spherical phantom with 18 cm diameter, this produced 4 slices across the field of view.

In Vivo Experiments

The SEAMS PINS sequence was used to scan the brain of a healthy human volunteer. Institutional Review Board approval and informed consent were obtained before scanning. The in vivo images were acquired with the same sequence parameters as the phantom images, but with a FOV of $240\text{ mm} \times 240\text{ mm}$. A 10.4 mT/m gradient was used with the slice thickness of 5 mm and a slice separation of 40 mm. To measure B_1 -insensitivity of SEAMS PINS, images were acquired with $\text{ODF} = 1.0, 1.1, 1.2, 1.4,$ and 1.6. The acquisition parameters for these scans are summarized in Table 1. The SNR was measured on the in vivo image as the mean signal acquired from 225 pixels divided by the standard deviation of the noise signal, acquired from 225 pixels in a nonimage region.

RESULTS

Figure 4 compares the data from the cylindrical water phantom, scanned with increasing RF power, with the simulated magnetization profiles for SEAMS PINS. The magnitude of the projection through the multislice profile for a range RF overdrive factors is plotted in Figure 4. Figure 4A shows the simulated slice profile and Figure 4B shows the measured slice profile in the

Table 1
Pulse Sequence Parameters for the Reference Scan and SEAMS PINS Image Acquisition for In Vivo and Phantom Scans

Sequence	TR [ms]	N_{av}	TE [ms]	FOV [mmxmm]	Matrix	Gradient [mT/m]	Thickness [mm]	Total time [min:sec]
Phantom reference	8.6	1	4	250x250	64x64	N/A	5	0:28
Phantom PINS	300	2	43	250x250	512x512	10.4	5	2:52
Human reference	8.6	1	4	240x240	64x64	N/A	5	0:27
Human PINS	300	4	43	240x240	256x256	10.4	5	5:20

cylindrical phantom experiment. In both figures, the slice parameters and position of the projection through the slices remain fixed. As expected, the amplitude of the selected slices increases as the angle of the excitation pulse approaches 90° and the spin echo refocusing pulse approaches adiabatic threshold (i.e., $ODF=1$). Past the adiabatic threshold for the refocusing pulse, magnetization declines at a much slower rate owing to the B_1 -insensitive behavior of the 180° pulse. Some loss of amplitude of the magnetization occurs due to the non-adiabatic behavior of the excitation pulse. Figure 4C compares the slice magnitude of a conventional windowed

sinc 90° – 180° spin echo pulse-pair with the simulated and experimental center slice magnitude in images produced by the SEAMS PINS pulse-pair, for a range of RF overdrive factors. Although there is some loss of slice signal as RF is overdriven above adiabatic threshold ($ODF=1$) due to the nonadiabatic behavior of the 90° pulse, the SEAMS PINS pulse-pair still achieves much slower signal decay than the conventional spin echo, due to the B_1 -immunity imparted by the adiabatic 180° PINS pulse.

Figure 5 illustrates the slice selection (profile of axial slices) of the SEAMS PINS sequence in a phantom. Figure 5A shows a coronal slice of a cylindrical water phantom. Figure 5B shows the multiple slice profiles obtained by using the SEAMS PINS sequence with the readout gradient in the slice-select direction. The slices shown in Figure 5B were 5 mm thick (full width at half maximum) and the separation between the center of each slice was 40 mm, as expected from the pulse design. The SEAMS PINS slice projection has been overlaid on the coronal water bottle slice image to provide a position reference. When the maximum gradient amplitude is decreased from 10.4 mT/m to 5.2 mT/m, the slice thickness and the slice separation increases to 10 mm and 80 mm, respectively (Fig. 5C), although the slice thickness to separation ratio remains constant at 1/8.

The B_1 -insensitivity of the SEAMS PINS sequence was explored through experiments on a spherical water phantom (Fig. 6). When the RF excitation and refocusing pulses were underpowered (Figure 6A; $ODF=0.5$), the un-aliased image quality is poor, as expected. Figure 6B illustrates an image acquired with the RF amplitude set to achieve adiabatic threshold at the center of the spherical phantom, while Figures 6C and 6D illustrate the images acquired with an increase in RF amplitude of 20% and 40%, respectively ($ODF=1.2$ and 1.4). As shown in Figure 6E, significant image degradation occurs with an ODF of 2.0 due to the nonadiabatic behavior of the 90° pulse and deterioration of phase-matching as the pulse-pair is overdriven (22). However, for a range of RF overdrive (up to 40%) images remain largely invariant, especially when compared with the center slice obtained through a conventional spin echo acquisition for the same range of ODF s (Figure 6, right column). This demonstrates the robust behavior of the SEAMS PINS pulse-pair in the presence of the B_1 variation typically observed in vivo at 7T. Figure 6F plots the central line projection through images obtained at the range of tested RF overdrive values. The semiadiabatic behavior of the SEAMS PINS sequence in the spherical phantom is consistent with simulated and measured values in the cylindrical water phantom.




FIG. 4. Behavior of simulated slice profiles and measured slice profiles in a phantom as RF amplitude is varied for SEAMS PINS. Simulation of the multislice pulse profile (A) and central cross-section of the slices (B) measured from imaging a cylindrical water phantom plotted for ODF s ranging from 0.5 to 1.5. C: Simulated and measured phantom slice profile amplitudes obtained using the SEAMS PINS sequence compared with simulated slice profile amplitudes obtained using a conventional nonadiabatic spin echo sequence.




FIG. 5. Slice profiles in cylindrical water phantom. **A**: The projection of the slice profile obtained by acquiring the readout direction parallel to the slice-select direction is overlaid on a conventionally acquired coronal image of the phantom. The slice profiles were acquired using gradient amplitudes of 10.4 mT/m, resulting in three 5 mm slices separated by 40 mm (**B**); and 5.2 mT/m, resulting in two 10-mm slices separated by 80 mm (**C**).

In vivo brain images obtained using the SEAMS PINS sequences are shown in Figure 7. The sum-of-squares images obtained by combining the image data from the 32 receive coils contains 5 overlapping slices (Figures 7A–E, left-most column) are shown for ODFs of 1, 1.1, 1.2, 1.4, and 1.6 in Figures 7A–E. The signal in the brain was measured in the solid white box showed in Figure 7A, and the noise measurement was taken from the region indicated by the dashed white box. The two measured values were used to compute the SNR. The same measurement, repeated in each image in the ODF experiment resulted in an SNR that ranged from a high of 48 (ODF=1.2), to a low of 34 (ODF=1.6). The SNR for ODF=1.0, 1.1., and 1.4 were 42, 46, and 43, respectively. The slices were disentangled using the low resolution reference scans (Fig. 7F), into 5 separate slices (Figures 7, remaining columns).

DISCUSSION

The SEAMS PINS sequence is capable of producing separable multislice spin echo profile with improved immunity to B_1 variation when compared with conventional spin echo sequences. Simulations show that for a conventional spin echo, a 20% percent increase in B_1 results in more than 10% signal loss; in the SEAMS sequence, an increase of 41% is required to see the same level of loss. Phantom data and in vivo experiments showed that the SEAMS PINS pulse sequence provides improved B_1 -insensitivity when compared with the traditional spin echo and resulted in minimal image quality variations over a 40% increase in transmit B_1 , making it a valuable pulse-pair for high-field sequences. Figure 4 illustrates the window of B_1 -insensitivity of the SEAMS PINS sequence. In Figures 6 and 7, phantom and in vivo images are shown over a 40% increase in RF amplitude above the adiabatic threshold. Note that B_1 insensitivity of transmission is demonstrated through the stability of the overdriven images in Figures 6 and 7. Image shading resulting from signal attenuation due to inhomogeneous B_1 reception profiles still exists; however this will only alter SNR, not contrast. When the B_1 amplitude is increased past an ODF of 1.4 to twice the adiabatic threshold (Fig. 6E), both the nonadiabatic behavior of the 90° pulse and loss of phase-matching create overall signal and contrast variation. In the in vivo experiment (Fig. 7), overall image quality was best when the center of the image was overdriven (ODF=1.2 and 1.4; Figures 7D and 7E) as compared to the case in which the center of the image was at adiabatic threshold (ODF=1.0; Figure 7A) but the off-center portions of the image were

underdriven below adiabatic threshold. However, when ODF=1.6 (Fig. 7E) the center of the image begins to lose signal as the pulse-pair amplitude exceeds the adiabatic range of behavior. Deterioration of phase-matching as RF amplitude increases beyond adiabatic threshold is an inherent disadvantage of the adiabatic matched-phase single-refocused spin echo approach when compared with an adiabatic twice-refocused spin echo (28). However, the use of a second identical refocusing pulse in a twice-refocused spin echo sequence would also significantly increase the minimum achievable TE and RF power deposited by the pulse sequence. Both shortening tissue T_2 values (12) and a quadratic increase in SAR with field strength necessitate the use of single-refocused, shorter TE alternatives for spin echo imaging at 7T. The matched-phase adiabatic pulse-pair provides a low-SAR alternative for achieving a spin echo with greater immunity to B_1 than conventional spin echo pulse-pairs.

Pulse BW and sampling rate, as well as gradient amplitude and duty cycle can be manipulated to adjust slice thickness and slice separation achieved by the SEAMS PINS pulse-pair. For the same gradient pulse shape and amplitude, decreasing the pulse BW alters the ratio of slice thickness to slice separation, resulting in thinner slices separated by the same distance. However, a minimum RF pulse BW must be maintained to achieve an adiabatic 180° pulse; thus there is a lower bound on how far the ratio can be adjusted using BW. The minimum slice thickness for our in vivo SEAMS PINS acquisition was gradient slew rate limited on our whole body scanner due to PNS safety limits. This is probably also true for head-only gradient systems with higher maximum slew-rates. Nevertheless, this hardware limitation exists for all RF methods using interleaved gradient blipping and RF sampling (i.e. 2D excitation, spectral-spatial RF pulses, etc). Given a required pulse BW and maximum gradient amplitude, thinner slices can either be obtained by reducing the RF sampling rate, resulting in a reduced slice separation, or by increasing the total duration of the pulse.

The SEAMS PINS sequence presented here uses a phase-matched 90° – 180° pulse-pair to achieve a minimum TE of 40 ms, including crushers, and gradient refocusing time. In comparison, an adiabatic PINS double echo with a PINS excitation would require a minimum TE of 53 ms, assuming comparable BW, crushers, gradient limits, and RF amplifier hardware limitations. Calculation of the SEAMS PINS SAR (SAR_{SEAMS}) and an adiabatic twice-refocused sequence ($SAR_{adiabatic}$) results in a relative SAR ratio ($SAR_{seams}/SAR_{adiabatic}$) of 0.56.




FIG. 6. Spherical water phantom images at different RF overdrive factors. Multiple overlapping slices obtained using SEAMS PINS (left column), four disentangled individual slices obtained from SEAMS PINS image (middle four columns), and single central slice obtained using a conventional spin echo sequence (right column) when RF pulse amplitude was set to 50% of adiabatic threshold ($ODF = 0.5$) (A); adiabatic threshold at the center of the sphere ($ODF = 1$) (B); 20% above adiabatic threshold ($ODF = 1.2$) (C); 40% above adiabatic threshold ($ODF = 1.4$) (D); and 100% above adiabatic threshold ($ODF = 2$) (E). F: Central cross-sections of slice 2 of the SEAMS PINS phantom images obtained at ODFs ranging from 0.5 to 2. These data show the robust behavior of the pulse-pair in the presence of moderate B_1 variation (typically observed over most of the cortex in vivo at 7T). However for severe changes in B_1 , as shown in F), pulse-pair behavior begins to degrade due to B_1 -sensitivity of the nonadiabatic 90° pulse and deterioration of phase-matching.

The TE could be shortened further by using a shorter 90° pulse which is still designed to unwind the phase of the 180° pulse. However, using this variant of the pulse-pair would result in degradation of the slice

profile in the presence of B_0 inhomogeneity. Shortening the 90° pulse can also be accomplished through shorter nulls. A consequence of this strategy is the need to increase the gradient slew rate maximum to achieve the




FIG. 7. Reconstructed multiple sagittal human brain images acquired using the SEAMS PINS sequence. Aliased slice (left) and individual un-aliased slices (right) for an image obtained at adiabatic threshold (A), ODF = 1.1 (B), ODF = 1.2 (C), ODF = 1.4 (D), and ODF = 1.6 (E). Dotted and solid white squares (in A, Slice 3) indicate ROIs used to measure SNR. F: Low-resolution reference images obtained to estimate the coil sensitivity profile.

same slice thickness, which may be limited by PNS constraints.

Successful separation of the multiple slices was demonstrated using a conventional readout. However, these pulses may be used to excite and separate more closely spaced slices by introducing interslice image shifts in the phase encoding direction using the CAIPIRINHA technique (29).

The pulse-pair may be integrated into any sequence that uses a single spin-echo. Diffusion MRI or spin echo functional MRI sequences are good candidates at high field strengths. For shorter TE, B_1 -insensitive multislice refocusing, a self-refocused adiabatic pulse as described in Balchandani et al (22) may also be modulated in the same way to generate a self-refocused adiabatic PINS pulse.

CONCLUSIONS

A SEAMS PINS pulse sequence using matched-phase adiabatic SLR pulses to generate a multislice spin echo was designed, implemented, and validated in phantoms and in vivo at 7T. Improved B_1 -immunity was demon-

strated by the SEAMS PINS sequence (10% signal loss over 41% increase in B_1) when compared with a traditional spin echo sequence (10% signal loss over a 20% increase in B_1). Quadratic phase deposited by the adiabatic 180° pulse was successfully refocused by the 90° pulse, obviating the need for a second adiabatic 180° pulse and therefore providing a 44% lower-SAR than a comparable twice-refocused adiabatic solution. Number of slices, slice thickness and separation are fully adjustable through changes in sampling frequency and/or gradient moment.

This technique may be applied in vivo to accelerate any spin echo sequence including those used for structural imaging, diffusion weighted imaging and functional imaging, while providing improved immunity to B_1 -inhomogeneity.

ACKNOWLEDGMENTS

We are grateful for the slice-GRAPPA Matlab script provided by Steen Moeller and Edward Auerbach from the Center for Magnetic Resonance Research (CMRR), University of Minnesota.

REFERENCES

- Ugurbil K, Adriany G, Andersen P, et al. Ultrahigh field magnetic resonance imaging and spectroscopy. *Magn Reson Imaging* 2003;21:1263–1281.
- Duyn JH. The future of ultra-high field MRI and fMRI for study of the human brain. *Neuroimage* 2012;62:1241–1248.
- Norris DG. High field human imaging. *J Magn Reson Imaging* 2003;18:519–529.
- Schick F. Whole-body MRI at high field: technical limits and clinical potential. *Eur Radiol* 2005;15:946–959.
- Heidemann RM, Porter DA, Anwander A, Feiweier T, Heberlein K, Knosche TR, Turner R. Diffusion imaging in humans at 7T using readout-segmented EPI and GRAPPA. *Magn Reson Med* 2010;64:9–14.
- Morelli JN, Runge VM, Feiweier T, Kirsch JE, Williams KW, Attenberger UI. Evaluation of a modified Stejskal-Tanner diffusion encoding scheme, permitting a marked reduction in TE, in diffusion-weighted imaging of stroke patients at 3 T. *Invest Radiol* 2010;45:29–35.
- Polders DL, Leemans A, Hendrikse J, Donahue MJ, Luijten PR, Hoogduin JM. Signal to noise ratio and uncertainty in diffusion tensor imaging at 1.5, 3.0, and 7.0 Tesla. *J Magn Reson Imaging* 2011;33:1456–1463.
- Stejskal EO, Tanner JE. Spin diffusion measurements: spin echoes in the presence of a time-dependent field gradient. *J Chem Phys* 1965;42:288–292.
- Duong TQ, Yacoub E, Adriany G, Hu X, Ugurbil K, Vaughan JT, Merkle H, Kim SG. High-resolution, spin-echo BOLD, and CBF fMRI at 4 and 7 T. *Magn Reson Med* 2002;48:589–593.
- Yacoub E, Duong TQ, Van De Moortele PF, Lindquist M, Adriany G, Kim SG, Ugurbil K, Hu X. Spin-echo fMRI in humans using high spatial resolutions and high magnetic fields. *Magn Reson Med* 2003;49:655–664.
- Vaughan JT, Garwood M, Collins CM, et al. 7T vs. 4T: RF power, homogeneity, and signal-to-noise comparison in head images. *Magn Reson Med* 2001;46:24–30.
- Cox EF, Gowland PA. Simultaneous quantification of T2 and T2* using a combined gradient echo-spin echo sequence at ultrahigh field. *Magn Reson Med* 2010;64:1440–1445.
- Peters AM, Brookes MJ, Hoogenraad FG, Gowland PA, Francis ST, Morris PG, Bowtell R. T2* measurements in human brain at 1.5, 3 and 7 T. *Magn Reson Imaging* 2007;25:748–753.
- Souza SP, Szumowski J, Dumoulin CL, Plewes DP, Glover G. SIMA: simultaneous multislice acquisition of MR images by Hadamard-encoded excitation. *J Comput Assist Tomogr* 1988;12:1026–1030.
- Weaver JB. Simultaneous multislice acquisition of MR images. *Magn Reson Med* 1988;8:275–284.
- Muller S. Simultaneous multislice imaging (SIMUSIM) for improved cardiac imaging. *Magn Reson Med* 1989;10:145–155.
- Feinberg DA, Setsompop K. Ultra-fast MRI of the human brain with simultaneous multi-slice imaging. *J Magn Reson* 2013;229:90–100.
- Larkman DJ, Hajnal JV, Herlihy AH, Coutts GA, Young IR, Ehnholm G. Use of multicoil arrays for separation of signal from multiple slices simultaneously excited. *J Magn Reson Imaging* 2001;13:313–317.
- Moeller S, Yacoub E, Olman CA, Auerbach E, Strupp J, Harel N, Ugurbil K. Multiband multislice GE-EPI at 7 tesla, with 16-fold acceleration using partial parallel imaging with application to high spatial and temporal whole-brain fMRI. *Magn Reson Med* 2010;63:1144–1153.
- Norris DG, Koopmans PJ, Boyacioglu R, Barth M. Power Independent of Number of Slices (PINS) radiofrequency pulses for low-power simultaneous multislice excitation. *Magn Reson Med* 2011;66:1234–1240.
- Koopmans PJ, Boyacioglu R, Barth M, Norris DG. Simultaneous multislice inversion contrast imaging using power independent of the number of slices (PINS) and delays alternating with nutation for tailored excitation (DANTE) radio frequency pulses. *Magn Reson Med* 2013;69:1670–1676.
- P, Khalighi MM, Glover G, Pauly J, Spielman D. Self-refocused adiabatic pulse for spin echo imaging at 7 T. *Magn Reson Med* 2012;67:1077–1085.
- Park JY, Garwood M. Spin-echo MRI using $\pi/2$ and π hyperbolic secant pulses. *Magn Reson Med* 2009;61:175–187.
- Balchandani P, Yamada M, Pauly J, Yang P, Spielman D. Self-refocused spatial-spectral pulse for positive contrast imaging of cells labeled with SPIO nanoparticles. *Magn Reson Med* 2009;62:183–192.
- Pauly J, Le Roux P, Nishimura D, Macovski A. Parameter relations for the Shinnar-Le Roux selective excitation pulse design algorithm [NMR imaging]. *IEEE Trans Med Imaging* 1991;10:53–65.
- Balchandani P, Pauly J, Spielman D. Designing adiabatic radio frequency pulses using the Shinnar-Le Roux algorithm. *Magn Reson Med* 2010;64:843–851.
- Setsompop K, Gagoski BA, Polimeni JR, Witzel T, Wedeen VJ, Wald LL. Blipped-controlled aliasing in parallel imaging for simultaneous multislice echo planar imaging with reduced g-factor penalty. *Magn Reson Med* 2012;67:1210–1224.
- Balchandani P, Qiu D. Semi-adiabatic Shinnar-Le Roux pulses and their application to diffusion tensor imaging of humans at 7T. *Magn Reson Imaging* 2014;32:804–812.
- Breuer FA, Blaimer M, Heidemann RM, Mueller MF, Griswold MA, Jakob PM. Controlled aliasing in parallel imaging results in higher acceleration (CAIPIRINHA) for multi-slice imaging. *Magn Reson Med* 2005;53:684–691.